

OPTICAL WAVELENGTH LABORATORIES

Customer Name Acme Corp. Date of Calibration 2/20/2012
Model Number KIT-FO4B-WSMDSDst Date of Next Calibration 2/20/2013
Transfer Standard ID FO40272 Calibrated by MG
Temp. (C) 24 Rel. Hum. (%) 34 Technician's Signature

NIST # 814342

OPTICAL LIGHT SOURCE CERTIFICATE OF CALIBRATION

Source Model Number WS-MDSDst
Source Serial Number WSA19999

Wavelength	850nm	1300nm	1310nm	1550nm
Coupled Power	-20.0 dBm	-20.0 dBm	-10.0 dBm	-10.0 dBm
Fiber Core Diameter	50 μ m	62.5 μ m	9 μ m	9 μ m
Connector Style	ST	ST	ST	ST

OPTICAL POWER METER CERTIFICATE OF CALIBRATION

Meter Model Number FO-4B
Meter Serial Number FO49999

Wavelength	850nm	1300nm	1310nm	1550nm
Transfer Standard	-10.0 dBm	-10.0 dBm	-10.0 dBm	-10.0 dBm
Before Calibration	dBm	dBm	dBm	dBm
After Calibration	-10.0 dBm	-10.0 dBm	-10.0 dBm	-10.0 dBm

Range Check	0 dBm	-10 dBm	-20 dBm
1310 nm	-30 dBm -29.98	-40 dBm -40.04	-20.02

N.I.S.T. TRACEABLE